



# CAPINTEC, INC.

## DECLARATION OF CONFORMITY THE EC DIRECTIVE AND CE MARKING

We, Capintec, Inc.  
620 Alpha Drive  
Pittsburgh, PA 15238  
Phone: 412-963-1988  
Fax: 412-963-0610

declare under our sole responsibility that the product CRC-ULTRA RADIOISOTOPE DOSE CALIBRATOR (including use with Okidata Microline 320, Epson Roll, Epson Slip, HP DeskJet 940c Inkjet and Epson Stylus C86 Inkjet printers) to which this declaration relates is in conformity with the Medical Device Directive 93/42/EEC.

This product is in conformity with the following standards or other normative documents following:

1. EN61010-1
2. EN60601-1-2

Certified to Annex V of the Directive 93/42/EEC on Medical Devices  
Dose Calibrators, Class I with measuring function  
Certified through Intertek Semko AB, Notified Body MDD  
CE Mark as shown denotes conformance to the above statement.



The Technical Construction File is maintained at:

Capintec, Inc. Phone: 412-963-1988  
620 Alpha Drive Fax: 412-963-0610  
Pittsburgh, PA 15238


The authorized representative located within the Community is:

Aries  
44 bis, Bd Felix Faure  
92323 Chatillion Cedex

Date of issue: 9/6/06

Place of issue:

Capintec, Inc.  
620 Alpha Drive  
Pittsburgh, PA 15238

  
Mary Anne Dell, Vice President & General Manager